

Extraction for Analysis

For analysis of extracted solution, please refer to analytical section of protocol manual.

Put 8 ml of pure water into extraction vial.



Slide sampler to remove from holder.



Remove sampler end-cap from sampler.



Remove top screen and then exposed filter from sampler using pincers.



Put filter into vial containing 8 ml of pure water.



Let stand for a period of 30 minutes, shaking several times during this period.

